

School Management Plan for Chait or Button

(Mic-Key or Mini-One)

that	
has been implanted into the abdomen to assist with personal hygiene. This device does not need to be accessed during school hours and should only be accessed at home.	
The following trouble shooting guide has been discussed with:	

Trouble shooting

The Chait/Button device is to assist with normal bowel function. The child may not wish to bring the device to the attention of other children. The child can play sport, play in the playground and swim with their device — please discuss this plan with the child's parent/carer first.

The Chait is dislodged and comes out of the abdomen a little, up to one (1) coil:

- This is not a medical emergency
- Reassure the child that this is ok
- · Gently push the Chait back in place if you are able
- Notify the child's parent/carer.

The Chait is pulled out of the abdomen:

- This is not a medical emergency
- Place the Chait in a plastic bag
- Notify the child's parent/carer who will bring the child to, or contact the Townsville University Hospital/local hospital, with the Chait as soon as possible.

The Button has dislodged from the abdomen:

- This is not a medical emergency
- Reassure the child that this is ok
- Notify the child's parent/carer for further instructions – they may have a spare with them that will need to be reinserted by an appropriately trained person.

There is some soiling on clothing near the Chait or Button:

- This is not a medical emergency
- Reassure the child that this is ok and see if they have a spare change of clothes
- Notify the parent/carer for further advice (the Button may have dislodged/Chait may have opened).

The child complains of pain or is knocked on or near the Chait/Button site:

 This may be a medical emergency. Assess the situation and call for an ambulance if needed.

Contact:	

If the child needs urgent medical attention, please call 000 or present to closest Emergency Department.

FACT SHEET: Children's Ward



Partnering with Consumers - This patient information brochure supports National Safety and Quality Health Service Standard 2. Consumers and/or carers provided feedback on this patient information.



